


Ultrasonic Cleaner

Date of revision: 24/09/24


Available in 5 Litre / 2.5 Litre / 1 Litre dispenser bottle

Intended Use / Intended Purpose: Enzyme based concentrate for the pre-cleaning and decontamination of the non-porous surfaces of hand instruments and forceps, rotary instruments, mirrors etc. Designed for use in the Ultrasonic bath. Specially formulated for use in the Medical and Dental environment.



1. Add 30ml ± 1ml of the concentrate to 970ml of tap water (3% solution).

The 1L bottle has a built in dosing device and the 2.5L a dosing cap.




2. In the ultrasonic bath completely immerse the instruments into the solution ensuring that all surfaces are covered by the solution, add hollow objects in such a way that air escapes.



3. Run the ultrasonic cleaner for a minimum of 15 minutes and according to the manufacturers instructions.



4. After immersion, remove the instruments and rinse them thoroughly prior to terminal sterilisation or high level disinfection.



5. Prepare a fresh solution daily or sooner if solution is very dirty. See product labels and MSDS for further information.

Composition: Aqua, Propan-2-ol, Didecyldimethylammonium chloride, Sodium n-lauroylsarcosinate, Subtilisin, Dye.

Storage. P210: Keep away from heat, hot surfaces, sparks, open flames and other ignition sources. No smoking. P403+P233: Store in a well-ventilated place. Keep container tightly closed. P235: Keep cool.

Contraindications. Aluminium, copper. Not recommended for plastics and rubbers. **Disposal.** Do not discharge concentrate into drains or rivers.

Intended Users. For professional use only. The products are used within the dental surgery by the dentist or dental assistant as part of their general disinfecting routines.

Patient Population: Dental surgery patients, both adults and children.

Any serious incident related to this medical device must be reported to the manufacturer and the competent authorities.

Warning. H226: Flammable liquid and vapour. H315: Causes skin irritation. H319: Causes serious eye irritation. H336: May cause drowsiness or dizziness. P302+350: IF ON SKIN: Gently wash with plenty of soap and water. P305+351+338: IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. P280: Wear protective gloves/protective clothing/eye protection/face protection. P261: Avoid breathing dust/fumes/gas/mist/vapours/spray. P304+340: IF INHALED: Remove person to fresh air and keep comfortable for breathing.



Warning.



For additional languages see www.topdental.co.uk



Manufactured in the UK by:
Topdental (Products) Ltd.
Unit 1, Holmfield Industrial Estate,
Halifax, W. Yorks, HX2 9TN. UK
Tel +44 (0)1535 652750
www.topdental.co.uk



DIRECTA AB P.O. Box 723,
Finvids väg 8 SE-194 27
Upplands Väsby, Sweden



UFI: 0610-E0JV-7002-VXNE